

Supplemental data Figure 1. Single Well Titer (SWT) of the different measurements (n=20) of the positive sample iQC (speckled and homogeneous) in the total imprecision experiment on the QUANTA-Lyser - NOVA View® system

The SWT are expressed as z-score with a z-score of 0 corresponding to a SWT of 160 (target value), z-score of ± 1 corresponding to a SWT of 80 and 320 and z-score of ± 2 corresponding to a SWT of <80 and 640.

